# 4452

# GENERAL PURPOSE MULTIPLY-DIVIDE-SQUARE-SQUARE ROOT OPERATOR

The Model 4452 is an economy multiplier/divider that requires only two external components to achieve maximum performance. Its small size and low cost make it an excellent choice for use as a computing element in the laboratory, in manufactured equipment, or wherever a multiplication process is required with no limitations on the polarity of input signals.

Unlike monolithic IC multipliers and most discrete multipliers, the Model 4452 requires no external amplifiers or circuitry, other than the two  $50k\Omega$  trimming potentiometers, for performing multiplication, division, squaring, or square-rooting. Selecting the mode of operation is determined by connecting the output of the module to the appropriate input pins.

The Model 4452 is fully encapsulated in epoxy for complete mechanical protection and for an almost completely isothermal environment for superior stability. The unit is short-circuit protected and the inputs are protected against overvoltage.

#### **APPLICATIONS**

As a multiplier there are no limitations on the polarity of input signals. Like other multipliers, when the 4452 is connected as a divider, the numerator, Z, can be either polarity, but the denominator, Y, must be positive and of such a magnitude that the output will not be required to exceed 10 volts in magnitude.




Figure 1A. Multiplication Mode





Figure 1B. Squaring Mode



#### **FEATURES**

- Low Cost
- Small Size
- High Input Impedance
- 4 Quadrant Operation
- No External Amplifiers Required

#### APPLICATIONS

- Automatic Gain Control
- Power Measurements
- Carrier Modulator/Demodulator
- Auto-Correlator
- Phase Detection

### Multiplication and Squaring Mode

- 1. Connect the Z terminal to the OUT terminal.
- Set X = 0 volts and Y = ±10 volts at 100 Hz. Adjust X<sub>OS</sub> for minimum output null as displayed on an oscilloscope.
- Set Y = 0 volts and X = ±10 volts at 100 Hz. Adjust Y<sub>OS</sub> for minimum output null as displayed on an oscilloscope.
- 4. For use as a squarer, connect terminal Y to terminal X.
- For use as a modulator, the carrier should be applied to the X terminal and the modulator to the Y terminal. Carrier null suppression is performed by adjusting the X<sub>OS</sub> potentiometer for dc offset, and the Y<sub>OS</sub> potentiometer for symmetry.

# SPECIFICATIONS Typical @ 25°C Vrc = ±15 V (unless noted otherwise)

SPECIFICATIONS Typical @ 25 C VCC - ±15 V (dilless noted offi	
TRANSFER FUNCTION Multiply Mode	$V_{out} = \frac{-x \cdot y}{10}$
Divide Mode	$V_{\text{out}} = \frac{-10 z}{y}$
Accuracy, % of Full Scale Trimmed	
4-Quadrant Operation	2% max,
2-Quadrant Operation	1% max.
INPUT	
Voltage	±10 V
Impedance X Input	40 k $\Omega$ min.

Z Input	90 kΩ
OUTPUT	
Voltage	±10 V
Current	±5 mA

Y Input

**FREQUENCY RESPONSE** -3 dB point 400 kHz 1° phase shift 2 kHz Full Output 40 kHz

STABILITY **Output Offset** vs Temperature **Output Noise** 

Impedance

3 mV/°C 2.5 mV RMS **TEMPERATURE RANGE** 

Operating Storage **POWER REQUIREMENTS** Supply Voltage

Current, Rated Output


Quiescent

±15 VDC ±1% ±15 mA ±5 mA

0°C to +70°C -25°C to +85°C

30 k $\Omega$  min. min.

 $1\Omega$ 



DIMENSIONS IN PARENTHESES ARE EXPRESSED IN CENTIMETERS

Optional Socket: NSK-20




Figure 2. Division Mode

#### **Division Mode**

- 1. Connect the X terminal to the OUT terminal.
- 2. Set Y = +10 volts and Z = 0 volts. Adjust  $X_{OS}$  for a minimum output null as measured on an oscilloscope.
- 3. Set Y = +10 volts and Z = +10 volts. Adjust  $Y_{OS}$  for -10volts output.

Note: Y must be positive when used as a divider, and of such value as will not require the output to exceed 10 V magnitude.




Figure 3. Square Root Mode

## Square Root Mode

- 1. Connect terminals X and Y to the OUT terminal.
- 2. Set both potentiometers to approximately mid range.
- 3. Set Z for -10 volts. Adjust the Yos potentiometer for an output of +10 volts.
- 4. Set Z for -1 volt. Adjust the X<sub>OS</sub> potentiometer for an output of +3.16 volts.
- 5. Repeat steps 3 and 4 until the required outputs have been obtained.
- 6. Voltage applied to Z should be between -1 volt and -10 volts.